

We, **X-CEN-TEK GmbH & Co. KG**, Westerburger Weg 30 - D-26203 Wardenburg, declare under our sole responsibility that the medical devices of the product group

## PAX Rescue & Transport Systems Basic UDI-DI: 426074548 016018020019 MJ

complies with the relevant provisions of Regulation (EU) 2017/745 (MDR) on medical devices and, where applicable, other relevant Union legislation

The product group includes the following medical devices						
Commercial name	Article No.	Commercial name	Article No.			
PAX Carrying Sheet basic	104490201	PAX Rescue Sheet with anti-slip protection L	102260201			
PAX Rescue Sheet XL with weight indication	102270201	PAX Carrying Sheet XXL 400 kg	101820201			
PAX Carrying Sheet light - tactical	277564503	PAX Carrying Sheet light - EMS	278890101			

Intended use of the product group: Rescue and transport of injured and sick persons

According to Annex VIII, Rule 1 MDR, all devices in the product group are class 1 medical devices and the applicable essential safety and performance requirements according to Annex I MDR are fulfilled. A conformity assessment procedure according to Annex II & III MDR has been carried out.

Applicable harmonized standards, national standards, or other regulatory documents: EN ISO 13485 – Medical devices – Quality management systems – Requirements for regulatory purposes EN ISO 14971 – Medical devices – Application of risk management to medical devices EN 1865-1 – Patient handling equipment used in road ambulances

This EU Declaration of Conformity is valid until **25<sup>th</sup> of May 2024** 

Nils-Lasse Schneider

Wardenburg, the 25<sup>th</sup> of May 2023

Dr. Nils-Lasse Schneider PRRC according to Art.15 MDR

## Manufacturers SRN: DE-MF-000009521

Version 1.0	BasDok erstellt: TC-05.04.2021	Freigabe QMB: TB-07.04.2021	Dok erstellt: TC	Freigabe VP: NLS-25.05.2023	
Datei: PAX CE	KE-EN TT 05-23		Anlage: 28.04.2021	Stand: 25.05.2023	3
Firma X-CEN-TEK GmbH & Co.KG – 26203 Wardenburg		QM-System nach EN ISO 13485		Seite 1 von 1	