

## EU - Declaration of Conformity



We, **X-CEN-TEK GmbH & Co. KG**, Westerburger Weg 30 - D-26203 Wardenburg, declare under our sole responsibility that the medical devices of the product group

### **PAX Rescue & Transport Systems**

**Basic UDI-DI: 426074548 016018020019 MJ**

complies with the relevant provisions of Regulation (EU) 2017/745 (MDR) on medical devices and, where applicable, other relevant Union legislation

#### The product group includes the following medical devices

Commercial name	Article No.	Commercial name	Article No.
<b>PAX Roll-Up Stretcher</b>	268935803	<b>PAX Vacuum pad - Roll-up</b>	268905310
<b>London Ambulance Service</b>		<b>Stretcher LAS</b>	

**Intended use of the product group:** Rescue and transport of injured and sick persons

According to Annex VIII, Rule 1 MDR, all devices in the product group are class 1 medical devices and the applicable essential safety and performance requirements according to Annex I MDR are fulfilled. A conformity assessment procedure according to Annex II & III MDR has been carried out.

**Applicable harmonized standards**, national standards, or other regulatory documents:

**EN ISO 13485** – Medical devices – Quality management systems – Requirements for regulatory purposes

**EN ISO 14971** – Medical devices – Application of risk management to medical devices

**EN 1865-1** – Patient handling equipment used in road ambulances

This EU Declaration of Conformity is valid until **25<sup>th</sup> of May 2024**

*Nils-Lasse Schneider*

Wardenburg, the 25<sup>th</sup> of May 2023

Dr. Nils-Lasse Schneider

PRRC according to Art.15 MDR

**Manufacturers SRN:** DE-MF-000009521

Version 1.0	BasDok erstellt: TC-05.04.2021	Freigabe QMB: TB-07.04.2021	Dok erstellt: TC	Freigabe VP: NLS-25.05.2023
Datei: PAX CE KE-EN RUS LAS 05-23			Anlage: 28.04.2021	Stand: 25.05.2023
Firma X-CEN-TEK GmbH & Co.KG – 26203 Wardenburg			QM-System nach EN ISO 13485	Seite 1 von 1