

We, **X-CEN-TEK GmbH & Co. KG**, Westerburger Weg 30 - D-26203 Wardenburg, declare under our sole responsibility that the medical devices of the product group

PAX Rescue & Transport Systems Basic UDI-DI: 426074548 016018020019 MJ

complies with the relevant provisions of Regulation (EU) 2017/745 (MDR) on medical devices and, where applicable, other relevant Union legislation

The product group includes the following medical devices					
Commercial name	Article No.	Commercial name	Article No.		
PAX RTS Cocoon - tactical	284430403				

Intended use of the product group: Rescue and transport of injured and sick persons

According to Annex VIII, Rule 1 MDR, all devices in the product group are class 1 medical devices and the applicable essential safety and performance requirements according to Annex I MDR are fulfilled. A conformity assessment procedure according to Annex II & III MDR has been carried out.

Applicable harmonized standards, national standards, or other regulatory documents:
EN ISO 13485 – Medical devices – Quality management systems – Requirements for regulatory purposes
EN ISO 14971 – Medical devices – Application of risk management to medical devices
EN 1865-1 – Patient handling equipment used in road ambulances

This EU Declaration of Conformity is valid until **25.05.2026**

Nils-Lasse Schneider

Wardenburg, 25.05.2025

Dr. Nils-Lasse Schneider PRRC according to Art.15 MDR

Manufacturers SRN: DE-MF-000009521

Version 1.1	BasDok erstellt: TC-05.04.2021	Freigabe QMB: TC-10.08.2023	Dok erstellt: TC	Freigabe VP: N	NLS- 25.05.2025
Datei: PAX CE KE-EN RTS-Cocoon 05-25.docx		Anlage: 26.10.2023	Stand: 25.05.2025		
Firma X-CEN-TEK GmbH & Co.KG – 26203 Wardenburg		QM-System nach EN ISO 13485		Seite 1 von 1	